

SYMBOL	SYMBOL TITLE	EXPLANATORY TEXT	STANDARD TITLE	STANDARD REFERENCE
	Manufacturer	Indicates the medical device manufacturer	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.1.1 ISO 7000 Reference No. 3082
XX REP	Authorized representative	Indicates the authorized representative in the identified country or jurisdiction The [XX] text of the symbol is replaced by either the two-letter country code or three-letter country code or other text required by the authority having jurisdiction. For example, EU REP means authorized representative in the European Union.	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.1.2
$\sim$	Date of manufacture	Indicates the date when the medical device was manufactured	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.1.3 ISO 7000 Reference No. 2497
$\Sigma$	Use-by date	Indicates the date after which the medical device is not to be used	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.1.4 ISO 7000 Reference No. 2607
LOT	Batch code	Indicates the manufacturer's batch code so that the batch or lot can be identified	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.1.5 ISO 7000 Reference No. 2492
REF	Catalogue number / Part number	Indicates the manufacturer's catalogue number/part number so that the medical device can be identified	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.1.6 ISO 7000 Reference No. 2493
SN	Serial number	Indicates the manufacturer's serial number so that a specific medical device can be identified	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.1.7 ISO 7000 Reference No. 2498
	Importer	Indicates the entity importing the medical device into the European Union	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.1.8 ISO 7000 Reference No. 3725
Ţ	Fragile, handle with care	Indicates a medical device that can be broken or damaged if not handled carefully	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.3.1 ISO 7000 Reference No. 0621
鯊	Keep away from sunlight and/or heat	Indicates a medical device that needs protection from light sources and/or heat	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.3.2 ISO 7000 Reference No. 0624
Ţ	Keep dry	Indicates a medical device that needs to be protected from moisture	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.3.4 ISO 7000 Reference No. 0626
	Temperature limit	Indicates the temperature limits to which the medical device can be safely exposed	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.3.7 ISO 7000 Reference No. 0632
(	Do not re-use / Single use	Indicates a medical device that is intended for one use, or for use on a single patient during a single procedure	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.4.2 ISO 7000 Reference No. 1051
ĺ	Consult instructions for use	Indicates the need for the user to consult the instructions for use	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.4.3 ISO 7000 Reference No. 1641
EIFU Indicator	Consult electric instructions for use	Indicates the need for the user to consult the instructions for use and where the electronic instructions for use (eIFU) can be found The eIFU indicator can be a manufacturer's website URL or some other appropriate indication that the instructions for use are available in an electronic format.	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.4.3 ISO 7000 Reference No. 1641



SYMBOL	SYMBOL TITLE	EXPLANATORY TEXT	STANDARD TITLE	STANDARD REFERENCE
LATEX	Contains natural rubber latex	Indicates natural rubber latex as a material of construction within the product	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.4.5 ISO 7000 Reference No. 2725
MD	Medical device	Indicates the item is a medical device	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.7.7 ISO 7000 Reference No. 2725
<b>A</b> →文	Translation	Indicates that the original medical device information has undergone a translation which supplements or replaces the original information	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.7.8 ISO 7000 Reference No. 3728
€ <sup>76</sup> ≥5	Repackaging	Indicates that a modification to the original medical device packaging configuration has occurred	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.7.9 ISO 7000 Reference No. 3727
UDI	Unique device identifier	Indicates a carrier that contains unique device identifier information This symbol may be used when multiple data carriers are present on the label. If used, this symbol is placed adjacent to the unique device identifier carrier.	Medical devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements	ISO 15223-1 Reference No. 5.7.10
CE	CE Mark	Indicates that a product complies with applicable European Union regulations	New approach directive Medical Devices Directive 93/42/EEC Medical Device Regulation (EU) 2017/745	New approach directive Medical Devices Directive 93/42/EEC Medical Device Regulation (EU) 2017/745
Ronly	Prescription Device	Indicates that the product is a medical device and Federal Law (USA) restricts this device to sale by or on the order of a dental professional	21 CFR 801.109 Prescription Devices	21 CFR 801.109 Prescription Devices